



राजस्थान मेडिकल एजुकेशन सोसायटी
(राजस्थान सरकार का उपक्रम)



कार्यालय, प्रधानाचार्य



राजमाता विजयाराजे सिंधिया मेडिकल कॉलेज एवं सम्बद्ध चिकित्सालय संघ, भीलवाड़ा

सिंदरी के बालाजी रोड, सांगानेर, भीलवाड़ा-311001(राजस्थान)

क्रमांक-RVRSMC/Purchase/2021/27

दिनांक:- 23/3/21

Request for Quotation

Kindly submit Quotation of following items for Microbiology Dept. Bhilwara as per following specifications:-

S.No.	Items	Quantity
1.	Fogger for Fumigation	-

Specifications of ULV Fogger for Fumigation:-

1. Motor: 1 HP with 1-3 rpm rotation, less time consuming and cost effective.
2. Power supply : 220 V , 50/60 Hz
3. Aerosols: The Fogger Machine provides excellent control of droplet size, even for near-invisible 0.5-10 micron droplets needed by concentrated ULV chemicals.
4. Discharge rate: 300 ml/min in a large (30 micron) droplets used in humidification, dust abatement and sanitizing work.
5. Space treatment up to 9000 cu feet & Above
6. Motor speed: 20000 rpm & above
7. Chemical level indicator, fogger can apply in either oil-or water –based chemicals.
8. Solution tank & lid: 4 Lt & above ; SS grade 304 & above.
9. Liquid filling nipple: SS 304, portable with SS tank set.

Note:- 1. Rates should include with GST.


2. Kindly sent quotations to the undersigned within 3 days in closed

Envelope with reference number of the request clearly mentioned on the envelope.

3. Incomplete Quotations (without reference number) will not be accepted.

4. Quotation sent on mail id will not be accepted.

5. FOR at Central store of college.


Principal & Controller
RVRS Medical College
Bhilwara.

