



राजस्थान मेडिकल एजुकेशन सोसायटी  
(राजस्थान सरकार का उपक्रम)

कार्यालय, प्रधानाचार्य



राजमाता विजयाराजे सिंधिया मेडिकल कॉलेज एवं सम्बद्ध चिकित्सालय संघ, भीलवाड़ा

सिंदरी के बालाजी रोड, सांगानेर, भीलवाड़ा-311001(राजस्थान)

क्रमांक-RVRSMC/Purchase/2021/ 38

दिनांक: 26/5/2021


### Request for Quotation

Kindly submit Quotation of following items for Microbiology Dept. Bhilwara Mycology Laboratory as per following specifications:-

S.N O.	ITEMS	Technical Specifications	Quantity
1	Analytic Chemical Balance	Capacity 120 gram, readability 0.1 mg, stabilization time 2 sec, SD= 0.1 mg, linearity +/-0.2 mg; ISO 9001, ISO 14001, WHO GMP certified with digital display	1
2	Water Bath	20 liters & above; temp range +50C-250 C, sensitivity +/-25°C, PT100 sensor. 1200-1500 W heater output: Auto off & on timer: Stainless steel tank SS 304 corrosion resistant and easy to clean, highly accurate PID controller with a bold 2 digit LED Display USA FDA class 1 medical device, accredited to ISO 9002	1
3	Digital Tissue Homogenizer	Hand operable with stainless steel probe for culture	1
4	Autoclave	<ol style="list-style-type: none"><li>1. Chamber Vol: 45Lts or above, top – Loading, foot pedal opening</li><li>2. Electrical Power: Voltage: 220 Volts 50 HZ</li><li>3. Internal heating elements within the autoclave chamber</li><li>4. Material of Construction: Inner &amp; outer chamber, steam generator made of corrosion-resistant stainless steel.</li><li>5. Temperature and pressure range 134°C or more, 4 bar</li><li>6. Number of Sterilization program built in up to 4, Available programs for can be individually parameter.</li><li>7. Code-secured access rights for changing parameters and further safety relevant intervention.</li><li>8. Sensor: Flexible PT-100 temperature sensor</li><li>9. Optional/Accessories: suitable stabilizer</li><li>10. Temperature holding function for liquids after program finish</li><li>11. It should be in fast cooling and reducing the cycle time for opening the door.</li><li>12. Should be FDA, CE Certified certificate and cGMP compliant, preferably ASME certified.</li><li>13. Temperature safety device preferably of class 2 &amp; 3.1 (DIN 12880) &amp; should meet UL/CSA/IEC safety standards.</li><li>14. Manufacturer should supply hard &amp; soft copies of user manual, technical, specifications.</li></ol>	1

Note:- 1. Rates should include with GST.

2. Kindly sent quotations to the undersigned within 3 days in closed Envelope with reference number of the request clearly mentioned on the envelope.
3. Incomplete Quotations (without reference number) will not be accepted.
4. Quotation sent on mail id will not be accepted.
5. FOR at Central store of college.

  
Principal & Controller  
RVRS Medical College  
Bhilwara

